

LC Systems

# Prominence-i Specifications

## *i*-Series Plus



# Specifications

		Prominence-i		
Model	LC-2030 ( UV model without sample cooler )	LC-2030C ( UV model )	LC-2030C 3D ( PDA model )	
P/N	228-65800-48	228-65801-48	228-65802-48	
Pump	Degassing unit	Five Lines: Mobile phase 4 + Rinse solution 1 (Volume 400 $\mu$ L)		
	Pumping method	Parallel-type double plunger		
	Pulsation	$\leq$ 0.1 MPa (1.0 mL/min, 10 MPa, Water)		
	Flow rate setting range	0.0001 to 10 mL/min		
	Flow rate accuracy	$\leq$ $\pm$ 1% or $\leq$ $\pm$ 2 $\mu$ L/min, below whichever is greater (0.01 to 2 mL/min, Specified condition) $\leq$ $\pm$ 2% (2 to 5 mL/min, Specified condition)		
	Flow rate precision	< 0.06 %RSD or < 0.02 minSD, below whichever is greater		
	Configuration	Four-solvent low-pressure gradient		
	Gradient / range of set concentrations	0 to 100%, in 0.1% steps		
	Gradient / concentration accuracy	$\pm$ 0.5% (0.1 to 2 mL/min, 1 to 20 MPa, Specified condition)		
	Gradient / concentration precision	$\pm$ 0.1% (1 mL/min, 10 MPa, Specified condition)		
	Maximum pressure	44 MPa (0.0001 to 5 mL/min), 22 MPa (5.0001 to 10 mL/min)		
	System Delay Volume	1100 $\mu$ L (Option: 650 $\mu$ L, 460 $\mu$ L)		
Autosampler	Injection method	Total-volume sample injection		
	Injection volume accuracy	$\pm$ 1% (50 $\mu$ L, N = 10)		
	Injection volume setting range	0.1 to 100 $\mu$ L (Option: 0.1 to 50 $\mu$ L, 1 to 500 $\mu$ L, 1 to 2,000 $\mu$ L)		
	Injection volume reproducibility	RSD < 0.20% (5.0–2000 $\mu$ L), RSD < 0.25% (2.0–4.9 $\mu$ L), RSD < 0.5% (1.0–1.9 $\mu$ L), RSD < 1.0% (0.5–0.9 $\mu$ L)		
	Cross-contamination	0.0025% (Caffeine, Specified condition)		
	Injection cycle time	Min. 14 sec (Specified condition)		
	Samples for processing	336 (1 mL), 216 (1.5 mL), 112 (4 mL), 4 (MTP/DWP)		
	Sample cooler	Not included	4 to 45°C (Room temperature needs to be 30°C or lower and the humidity 70% or less to go down to 4°C. Room temperature needs to be 15°C or higher to go up to 45°C.)	
	Injection linearity	> 0.9999 (1 to 100 $\mu$ L, Specified condition)		
	Pretreatment function	Auto-dilution, Reagent addition, Co-injection		
Column Oven	Heating and cooling method	Forced air circulation method		
	Containable column size	6 pieces at 10 cm max., 3 pieces at 10 cm to 30 cm		
	Temperature control range	Room temperature – 12 to 90°C, Setting range 4 to 90°C		
	Temperature control precision	$\pm$ 0.1°C		
	Temperature stability	$\pm$ 0.8°C (Specified condition)		
	Flow rate switching valve	Max. 1 pc		

		Prominence-i	
Model	LC-2030 (UV model without sample cooler)	LC-2030C (UV model)	
P/N	228-65800-48	228-65801-48	
UV Detector	Wavelength range	190 to 700 nm	
	Spectrum slit width	8 nm	
	Wavelength accuracy	≤ ±1 nm	
	Wavelength reproducibility	≤ ±0.1 nm	
	Noise level	≤ ±2.5 × 10 <sup>-6</sup> AU, (250 nm, Specified condition)	
	Drift	≤ 100 × 10 <sup>-6</sup> AU/h (250 nm, Specified condition)	
	Simultaneous monitoring of 2 wavelengths	Enable (Any 2 wavelengths of 190 to 370 nm or 371 to 700 nm)	
	Linearity	Up to 2.5 AU (5%)	
	Sampling rate	Up to 100 Hz	
	Light source	Deuterium (D2) lamp	
	Flow cell	12 µL (10 mm, TC), 12 MPa	
	Option cell	High-Speed: 8 µL (10 mm, TC), Semi-micro: 2.5 µL (5mm, TC)	

		Prominence-i	
Model	LC-2030C 3D (PDA model)		
P/N	228-65802-48		
PDA Detector	Wavelength range	190 to 800 nm	
	Spectral resolution	1.4 nm (Specified condition)	
	Slit width	1.2 nm, 8 nm	
	Device resolution	0.6 nm/pixel	
	Number of photodiode array elements	1024	
	Wavelength accuracy	≤ ±1 nm	
	Noise level	≤ ±3 × 10 <sup>-6</sup> AU (250 nm, reference: 350 nm, Specified condition)	
	Drift	≤ 500 × 10 <sup>-6</sup> AU/h (250 nm, reference: 350 nm, Specified condition)	
	Linearity	Up to 2 AU (5%)	
	Sampling rate	Up to 100 Hz	
	Light source	Deuterium (D2) lamp (Standard), tungsten (W) lamp (option)	
	Flow cell	10 µL (10mm, TC) 12 MPa	
Option cell	High-Speed: 8 µL (10 mm, TC), Semi-micro: 2.5 µL (5 mm, TC)		

		Prominence-i		
Model	LC-2030 ( UV model without sample cooler )	LC-2030C ( UV model )	LC-2030C 3D ( PDA model )	
P/N	228-65800-48	228-65801-48	228-65802-48	
Miscellaneous	Dimensions	W410×H605×D500 mm		
	Weight	58 kg	63 kg	
	Available pH range	1 to 13		
	Materials for parts in contact with liquids	Stainless steel (SUS316L, SUS316), FEP, PEEK, PTFE, perfluoroelastomer, ruby, sapphire, Hastelloy C, GFP, ceramic, PFA, quartz, PPS		
	Workstation	LabSolutions LC/GC Ver.5.71 or later, LabSolutions DB/CS Ver.6.31 or later (Incompatible with LCsolution)		

## Main Optional Accessories

### Solvent Delivery Units

Part Name	P/N	Description
FCV-11AL	228-45048-58	This is the mobile phase selection valve (3 flow lines). An FCV-11AL connection kit is required to connect to an FCV-11AL unit.
FCV-11ALS	228-45049-91	This is the mobile phase selection valve (1 flow line). An FCV-11AL connection kit is required to connect to an FCV-11AL unit.
FCV-11AL Connection Kit	228-56249-41	This kit includes connector cables and other items necessary for connecting FCV-11AL and FCV-11ALS units.
780 µL Mixer Kit	228-57313-41	This parts set includes a mixer and tubing for using TFA or other UV-absorbing substance as a mobile phase.
2 mL Mixer Kit	228-57313-42	This parts set includes a mixer and tubing for using TFA or other UV-absorbing substance as a mobile phase.
Compatible Volume System Kit	228-57796-41	This kit decreases the system volume to 650 µL.
Low Volume System Kit	228-57796-42	This kit decreases the system volume to 460 µL.

### Autosamplers

Part Name	P/N	Description
100 µL Sample Loop	228-56074-42	This sample loop is used for injecting 100 µL volumes. (Standard configuration parts of Prominence-i)
Optional 500 µL Sample Loop	228-45405-41	This increases the injection volume to 500 µL.
Optional 2 mL Sample Loop	228-45405-42	This increases the injection volume to 2 mL.
UHPLC Fitting (set of 1)	228-56867-41	Fitting for inlet to high-pressure capacity column
UHPLC Fitting (set of 10)	228-56867-43	Fitting for inlet to high-pressure capacity column
Sample Rack	228-55735-41	Additional sample rack
Plate for 1 mL Sample Vials (set of 2)	228-56197-41	Plate used to place 84 1 mL sample vials
Plate for 1.5 mL Sample Vials (set of 2)	228-50830-92	Plate used to place 54 1.5 mL sample vials
Plate for 4 mL Sample Vials (set of 2)	228-56197-42	Plate used to place 28 4 mL sample vials
Metal plate for 1.5ml Sample Vials (set of 1)	228-61515-42	Plate used to place 54 1.5 mL sample vials

### Column Ovens

Part Name	P/N	Description
Column Clamp ASSY B5	228-15617-91	This set of clamps is for adding a column with an outside diameter between 6.4 and 9.5 mm.
Column Clamp ASSY B8	228-15617-92	This set of clamps is for adding a column with an outside diameter between 9.5 and 12.7 mm.
FCV-14AH	228-45014-41	Automatic column switching valve with 6 positions and 7 ports which is usable at a pressure of 34.3 MPa max.
FCV-34AH	228-45185-41	Automatic column switching valve with 6 positions and 7 ports which is usable at a pressure of 100 MPa max.
FCV Mounting Kit	228-55765-42	This parts kit is used to secure an FCV-14AH/ 34AH unit inside the column oven.
CMD	228-37281-41	This column management device is used to record information about columns.
CMD Cable	228-39991	This cable is used to connect between the CMD and main units.

### UV Detectors

Part Name	P/N	Description
Recycle Valve	228-56808-41	This low-pressure flow-line selection valve is used to recycle mobile phase.
Flow Cell for UV Detectors	228-56167-41	This cell is compatible with conventional analysis. (Standard configuration parts of Prominence-i)
Semi-Micro Cell for UV Detectors	228-45605-46	This cell is compatible with semi-micro analysis.

### PDA Detectors

Part Name	P/N	Description
W Lamp ASSY for PDA Detectors	228-57110-41	This assembly includes a tungsten lamp and its socket used for high-sensitivity analysis in the long-wavelength region.
Flow Cell for PDA Detectors	228-42593-43	This cell is compatible with conventional analysis. (Standard configuration parts of Prominence-i)
High-Speed Cell for PDA Detectors	228-45618-54	This cell is compatible with fast analysis. (Standard configuration parts of Nexera-i)
Semi-Micro Cell for PDA Detectors	228-45605-47	This cell is compatible with semi-micro analysis.

## Other Options

Part Name	P/N	Description
Earthquake Reinforcement Kit	228-56298-41	This kit is used to reinforce how the reservoir tray is attached.
1 L Mobile Phase Bottles (set of 5)	228-38583-42	This is a set of five one-liter reservoir bottles for holding mobile phases.
Optional Detector Attachment Kit	228-56245-41	This kit contains a top plate and reservoir tray for installing an additional detector.
Optional Optical Board	228-55518-41	This board is used to install additional connectors for optical link cables. It is used to install fluorescence detector RF-20A series and other detectors.
Camera ASSY for Autosampler	228-55517-41	This camera is installed inside autosamplers. It allows you to monitor the needle action via the computer screen.
Optional AD Board	228-55519-41	This is an analog-digital converter board. It is used to input the detector signal as an analog signal, such as when a non-Shimadzu detector is connected.
Touch Panel Protecting Sheet	228-59212-41	Protecting sheet for touch panel.
Upgrade Kit UV	228-66139-43	Kit for upgrade from Prominence-i (UV model with sample cooler) to Nexera-i Plus.
Upgrade Kit PDA	228-66139-44	Kit for upgrade from Prominence-i (PDA model) to Nexera-i Plus.
Smart Automation Kit (4-mobile phase)	228-26004-44	This kit includes FCV-14AH for up to six columns switching and other parts.
Smart Automation Kit (7-mobile phase)	228-26004-43	This kit includes FCV-14AH for up to six columns switching, FCV-11AL for solvent delivery of seven mobile phases and other parts.

## Optional Detector Specifications



### RID-20A

	RID-20A (228-45104-XX)
Reflective index measurement range	1 to 1.75 RIU
Noise level	≤ 2.5 nRIU
Drift	≤ 0.1 μRIU/h
Range	A mode: 0.01 to 500 μRIU P and L modes: 1 to 5000 μRIU
Response	No filtering, 0.05 to 10 sec, 11 steps
Polarity switching	With a switch
Zero adjustment	Auto zero, auto optical zero, baseline shift functions
Maximum operating flow rate	20 mL/min (150 mL/min with an option)
Temperature control of cell unit	30 to 60°C (0.01°C steps)
Cell capacity	9 μL
Material in contact with liquid	SUS316L, quartz, PTFE, Al <sub>2</sub> O <sub>3</sub> , ETFE
Maximum operating pressure	0.4 MPa (4 kgf/cm <sup>2</sup> )
Operating temperature range	4 to 35°C
Dimensions and weight	W260 × D420 × H140 mm, 12 kg

Note: Hexafluoroisopropanol (HFIP) cannot be used as the mobile phase.



### RF-20A/RF-20Axs

	RF-20A (228-45147-XX)	RF-20Axs (228-45148-XX)
Light source	Xenon lamp	Xenon lamp, low-pressure mercury lamp (To check wavelength accuracy)
Wavelength range	0, 200 to 650 nm	0, 200 to 750 nm
Spectral bandwidth	20 nm	
Wavelength accuracy	±2 nm	
Wavelength precision	±0.2 nm	
S/N	Water Raman peak S/N 1200 min. Low background S/N > 9000	Water Raman peak S/N 2000 min. Low background S/N > 12000
Cell capacity	12 μL, 2 MPa (approx. 20 kgf/cm <sup>2</sup> ), SUS316L, PTFE (fluororesin), quartz	
Cell temperature control range	—	4 to 40°C, 1°C steps
Cell temperature setting range	—	(Room temperature – 10°C) to 40°C
Functions	Four-wavelength detection, wavelength scanning	
Safety measures	Liquid-leakage sensor	
Operating temperature range	4 to 35°C	
Dimensions and weight	W260 × D420 × H210 mm, 16 kg	W260 × D420 × H210 mm, 18 kg



Shimadzu Corporation

[www.shimadzu.com/an/](http://www.shimadzu.com/an/)

**For Research Use Only. Not for use in diagnostic procedures.**

This publication may contain references to products that are not available in your country. Please contact us to check the availability of these products in your country.

Company names, products/service names and logos used in this publication are trademarks and trade names of Shimadzu Corporation, its subsidiaries or its affiliates, whether or not they are used with trademark symbol "TM" or "®".

Third-party trademarks and trade names may be used in this publication to refer to either the entities or their products/services, whether or not they are used with trademark symbol "TM" or "®".

Shimadzu disclaims any proprietary interest in trademarks and trade names other than its own.

Shimadzu does not assume any responsibility or liability for any damage, whether direct or indirect, relating to the use of this publication.